

JAPANESE INDUSTRIAL STANDARD

JIS T 9001 : 2021

Performance requirements and test methods of medical face masks and non-medical face masks

October, 2023

ERRATUM

Page	Position	Error	Correct
19	5.1.4.3 h)	...a metal ring for retaining a test piece, 75 mm in inside diameter, 110 mm in outside diameter and 5 mm in thickness...	...a metal ring for retaining a test piece, 79 mm in inside diameter, 110 mm in outside diameter and 5 mm in thickness...
