



JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS B 7529 : 2017

(JMIF/JPTMA/JSA)

Vapour pressure thermometers

ICS 17.200.20

Reference number : **JIS B 7529 : 2017 (E)**

Date of Establishment: 1963-11-01

Date of Revision: 2017-08-21

Date of Public Notice in Official Gazette: 2017-08-21

Investigated by: Japanese Industrial Standards Committee
Standards Board for ISO area
Technical Committee on Basic Engineering

JIS B 7529:2017, First English edition published in 2018-01

Translated and published by: Japanese Standards Association
Mita MT Building, 3-13-12, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2018

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

AT

Contents

	Page
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Names of major parts	3
5 Classification and symbol for designation	3
6 Scale marking	4
7 Appearance and configuration	6
8 Shapes and dimensions	6
8.1 Principal dimensions of indicator	6
8.2 Shapes and dimensions of connecting part	8
8.3 Shapes and dimensions of temperature sensitive part	9
9 Performance and test methods	10
9.1 General	10
9.2 Test conditions	10
9.3 Performance	11
9.4 Performance test	11
9.5 Resistance to high temperature	11
9.6 Step response property	11
9.7 Resistance to vibration	12
9.8 Position of use	12
10 Inspection	12
10.1 Classification of inspections	12
10.2 Type inspection	12
10.3 Delivery inspection	12
11 Designation of product	13
12 Marking	13

Foreword

This translation has been made based on the original Japanese Industrial Standard revised by the Minister of Economy, Trade and Industry through deliberations at the Japanese Industrial Standards Committee as the result of proposal for revision of Japanese Industrial Standard submitted by Japan Measuring Instruments Federation (JMIF)/Japan Pressure Gauges and Thermometers Manufacturers' Association (JPTMA)/Japanese Standards Association (JSA) with the draft being attached, based on the provision of Article 12 Clause 1 of the Industrial Standardization Law applicable to the case of revision by the provision of Article 14.

Consequently **JIS B 7529**:1979 is replaced with this Standard.

This **JIS** document is protected by the Copyright Law.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, applications for a patent after opening to the public or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, applications for a patent after opening to the public or utility model rights.

Vapour pressure thermometers

1 Scope

This Japanese Industrial Standard specifies the vapour pressure thermometers (hereafter referred to as thermometers) consisting of a Bourdon tube and a temperature sensitive cylinder connected with each other by a conduit tube, and filled with a volatile liquid. The indicator in a round case has a single pointer and scales concentrically marked on the dial. The pressure change caused by the vapour pressure change of the volatile liquid contained in the system according to the temperature change at the temperature sensitive part is converted to a mechanical displacement of the pointer to indicate the temperature. The thermometers are used in the temperature range of $-30\text{ }^{\circ}\text{C}$ to $+200\text{ }^{\circ}\text{C}$.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this Standard. For those without the indication of the year, the most recent edition (including amendments) shall be applied.

JIS B 0202 *Parallel pipe threads*

JIS B 0203 *Taper pipe threads*

JIS Z 8103 *Glossary of terms used in measurement*

JIS Z 8703 *Standard atmospheric conditions for testing*

3 Terms and definitions

For the purpose of this Standard, the following terms and definitions, and those given in **JIS Z 8103** apply.

3.1 pressure system

system for converting a measured temperature into a visible displacement, consisting of a Bourdon tube, a temperature sensitive cylinder and a conduit tube, which is filled with a volatile liquid

3.2 indicator

part consisting of a Bourdon tube, an amplification mechanism connected with a temperature sensitive part via a conduit tube, and a case for housing those parts, which indicates a temperature

3.3 conduit tube

metal capillary connecting a temperature sensitive cylinder with a Bourdon tube

3.4 Bourdon tube

flattened metal tube formed into an arc and connected to the temperature sensitive cylinder via the conduit tube, which generates the displacement by vapour pressure change due to temperature of the volatile liquid contained in the temperature sensitive part