

Translated and Published by Japanese Standards Association

## $JIS \ T \ 3215^{\,:\,2021}$

### (MTJAPAN/JSA)

# Drainage catheters and accessory devices

ICS 11.040.25 Reference number : JIS T 3215 : 2021 (E)

Date of Establishment: 2005-03-25

Date of Revision: 2021-03-01

Date of Public Notice in Official Gazette: 2021-03-01

Investigated by: Japanese Industrial Standards Committee

Standards Board for ISO area

Technical Committee on Medical Equipment

JIS T 3215 : 2021, First English edition published in 2022-02

Translated and published by: Japanese Standards Association Mita MT Building, 3-13-12, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

> In the event of any doubts arising as to the contents, the original JIS is to be the final authority.

© JSA 2022

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

#### Contents

#### Page

Introduction ······1		
1	Scope1	
2	Normative references ······2	
3	Terms and definitions ······2	
4	Configuration an	nd names of parts ······3
$5 \\ 5.1 \\ 5.2$	Requirements 4   General requirements 4   Specific requirements 5	
6 6.1 6.2	Packaging ······ 6 Primary packaging ····· 6 Secondary packaging ····· 7	
7 7.1 7.2 7.3	Marking7Primary packaging7Secondary packaging7Use of symbols8	
Annex A (informative) Test method for determining kink stability9		
Annex	x B (normative)	Test method for corrosion resistance11
Annex	c C (normative)	Test method for resistance to deformation by suction $\cdots \cdots 12$
Annex	D (normative)	Test method for determining peak tensile force of con- nections ······13
Annex	E (normative)	Test method for determining peak tensile force of drainage catheter14
Annex	x F (normative)	Test method for impact resistance of collection device16
Annex	G (normative)	Test method for resistance to leakage when negative pressure is applied
Annex JA (informative) Comparison table between JIS and corresponding International Standard		

#### Foreword

This Japanese Industrial Standard has been revised by the Minister of Health, Labour and Welfare through deliberations at the Japanese Industrial Standards Committee as the result of proposal for revision of Japanese Industrial Standard submitted by Medical Technology Association of Japan (MTJAPAN)/Japanese Standards Association (JSA) with a draft being attached, based on the provision of Article 12, paragraph (1) of the Industrial Standardization Act applied mutatis mutandis pursuant to the provision of Article 16 of the said Act. This edition replaces the previous edition (**JIS T 3215** : 2011), which has been technically revised.

However, **JIS T 3215** : 2011 remains valid for three years from the date of public notice of the revision of this Standard.

This **JIS** document is protected by the Copyright Act.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, published patent application or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, published patent application or utility model rights.

#### Drainage catheters and accessory devices

#### Introduction

This Japanese Industrial Standard has been prepared based on **ISO 20697** : 2018, Edition 1, with some modifications of the technical contents in order to reflect the actual situation in Japan.

The vertical lines on both sides and dotted underlines indicate changes from the corresponding International Standard. A list of modifications with the explanations is given in Annex JA.

#### 1 Scope

This Standard specifies requirements for drainage catheters and their accessory devices, where the catheter is placed in a body cavity or wound, surgically or percutane ously, for drainage of accumulated fluid or air to the exterior. This Standard is not applicable to those with a special shape at the tip (pigtail, balloon, etc.) to maintain placement in the body, or to the following :

- suction catheters used by airway;
- tracheal catheters (tracheal tube);
- urethral catheter;
- ureteral stents, biliary stents, and other stents;
- drainage catheters placed in digestive tracts percutaneously with gastrostomy technique;
- neuraxial catheters used for removal of cerebrospinal fluid;
- enteral catheters used for removal of solutions or substances from the gastrointestinal tract;
- catheters for nephrostomy or cystostomy;
- catheters and tubes designed for the biliary tract.
  - NOTE The International Standards corresponding to this Standard and the symbol of degree of correspondence are as follows.

**ISO 20697** : 2018 Sterile drainage catheters and accessory devices for single use (MOD)

In addition, symbols which denote the degree of correspondence in the contents between the relevant International Standards and **JIS** are IDT (identical), MOD (modified), and NEQ (not equivalent) according to **ISO/IEC Guide 21-1**.